



FACSCalibur (Becton Dickinson) - 4 colours Software: Cell Quest

ModFit LT

Performance

Fluorescence Sensitivity	Estimated detection limit is 750 molecules of equivalent soluble fluorescein
Fluorescence Resolution	Coefficient of variation in FL2-Area of < 3%, full peak for propidium iodide-stained chicken erythrocyte nuclei
Forward and Side Scatter	Sensitivity enables the separation of fixed platelets from
Sensitivity	noise
Forward and Side Scatter	Scatter performance is optimized for resolving
Resolution	lymphocytes, monocytes and granulocytes

Fluidics

General operation	Front panel control with three modes: RUN, STNDBY, and PRIME; standby mode for conserving sheath fluid when no sample tube installed
Fluid reservoirs	2 Lt sheath and waste containers
Sheath pressure	4.5 psi
Sample flow rates	Regulated and monitored pressure difference between sheath and sample; particle velocity in flow cell approximately 6 meters/second Low = 12 µL/min Medium= 35 µL/min High = 60 µL/min
Maximum acquisition rate	2,000 events/sec
Recommended sample concentration range	Single-cell suspension of 1 x 10 ⁵ to 2 x 10 ⁷ particles/mL

T+351 239 820 190

F+351 239 822 776

Azinhaga de Santa Connar, 3004-504 Coimbra, Portugal

T+351 239 480 200

Universidade de CoimbraPólo III – Pólo das Ciências da SaúdeUC – Biotech,Rua Larga, Faculdade de Medicina,
Pólo I, 1º andarUniversidade de CoimbraPárque Tecnológico de Cantanhede3004-504 Coimbra, Portugal3004-504 Coimbra, PortugalNúcleo 04, Lote 81 T+351 231 249 170



Laser	Detector	Filter (nm)	Fluorochrome
Blue (488 nm)	FL1	BP 530/30	FITC, GFP, CFSE, DCFH
	FL2	BP 585/40	PE , PI, PE-Texas Red
	FL3	LP 670	PerCP-Cy5.5, PE-Cy5.5, PE-Cy7, PI, 7-AAD
Red (635nm)	FL4	BP 661/16	APC , Cy5, TOTO-3, Alexa 635

Emission Optics

Collection lens	Optical gel-coupled to quartz cuvette
Fluorescence detection	4 PMT detectors (Standard) + 1 optional 3 wavelengths detected from 488-nm laser: 530/30 nm 585/42 nm >670 nm One wavelength detected from optional 633-nm laser: 661/16 nm
Forward scatter detection	Photodiode with 488/10 bandpass filter
Side scatter detection	PMT with Brewster-angle beam splitter

Optical Layout

The following figure show a schematic diagram of the FACSCalibur optical layout, with detectors (FSC, SSC, FL1-4) and example fluorochromes that can be used when designing experiments for this instrument.

The following figure shows a schematic representation of the FACSCalibur optical layout.

> Universidade de Coimbra Rua Larga, Faculdade de Medicina, Universidade de Coimbra Pólo I, 1º andar 3004-504 Coimbra, Portugal

T+351 239 820 190

F+351 239 822 776

3004-504 Coimbra, Portugal


T+351 239 480 200

 Pólo III – Pólo das Ciências da Saúde
 UC – Biotech,

 Universidade de Coimbra
 Parque Tecnológico de Cantanhede

 Azinhaga de Santa Comba, Celas,
 Núcleo 04, Lote 8

 3060-197 Cantanhede, Portugal T+351 231 249 170



Universidade de Coimbra Rua Larga, Faculdade de Medicina, Pólo I, 1º andar 3004-504 Coimbra, Portugal

T+351 239 820 190

F+351 239 822 776

Pólo III – Pólo das Ciências da Saúde Universidade de Coimbra Azinhaga de Santa Comba, Celas, 3004-504 Coimbra, Portugal

T+351 239 480 200

UC – Biotech, Parque Tecnológico de Cantanhede Núcleo 04, Lote 8 Nucleo U4, Lote 8 3 3060-197 Cantanhede, Portugal

info@cnc.uc.pt www.cnc.uc.pt

T+351 231 249 170



Partec CyFlow[®] space - 7 colours

Software: FloMax®

Performance

Fluorescence Sensitivity	Superior fluorescence sensitivity of < 100 MESF (FITC) and < 50 MESF (PE)
Fluorescence Resolution	Coefficient of variation of about 2% on all fluorescence channels
Forward and Side Scatter Resolution	Going to nano-resolution in the scatter channel

Fluidics

Flow Cuvette	Synthetic quartz flow cuvette (350 x 200 μm) for laminar sample transport with sheath fluid for fluorescence, forward and side scatter light detection
Sample Delivery	Computer controlled precision syringe pump for contamination-free sample transport. Built-in air pressure for sheath fluid. Sheath fluid pressure is adjustable from 0- 300 mbar (computer controlled). Default setting: 200 mbar
Sampling Volume	Continuous up to 1500 µl; 200 µl for precise absolute counting
Flow Rates	Sample volume speed adjustable continuously between 0 and 50 µL/s. Sheath fluid pressure continuously adjustable
Maximum acquisition rate	52,000 events/sec
Fluidics Volume	2 Lt reservoirs for sheath fluid and waste
BioSafety System	Avoids sample droplets and sample cross contamination (computer controlled)

T+351 239 820 190

F+351 239 822 776

T+351 239 480 200

Universidade de CoimbraPólo III – Pólo das Ciências da SaúdeUC – Biotech,Rua Larga, Faculdade de Medicina,Universidade de CoimbraParque Tecnológico de CantanhedePólo I, 1º andarAzinhaga de Santa Comba, Celas,Núcleo 04, Lote 82004-504 Coimbra Portugal3004-504 Coimbra, Portugal3060-197 Cantanhede, Portugal T+351 231 249 170



Laser	Detector	Filter (nm)	Fluorochrome
Blue (488 nm)	FL1	BP 536/40	FITC, GFP, Alexa Fluor 488, Syto 9
	FL2	BP 590/50	PE, YFP
	FL3	BP 675/10	PE-Texas Red
	FL4	LP 748	PE-Cy5
Red (635nm)	FL5	BP 675/10	APC , Cy5, TOTO-3, Alexa 635
	FL6	LP 748	APC-Cy7
Violet (405 nm)	FL7	BP 455/25	Alexa Fluor 405, Pacific Blue

Emission Optics

Fluorescence detection	Nine optical parameters, each with its own photomultiplier tube (PMT), including forward scatter PMT and side scatter PMT
Optical parameters	FSC, SSC, FL1-FL7

The following figure shows an example selection of fluorochromes for the CyFlow® space Light source system.

Type of laser		Fluorescen	ce Channel	Flu	orochromes				
blue solid state laser 48 200 mW	88 nm (ORANGE RED RED	FITC PE PE-Texas F PE-Cy5 PE-Cy7	GFP eYFP Red® PE-Cy5.5	Alexa Fluor ⁹ Pl PerCP	PerCP-Cy5.5	SYTO 9 ECD	EB
red diode laser 63 25 mW	35 nm	0	RED INFRA RED	APC APC-Cy7	APC-Cy5.5 SYTO 59-6	61 Alexa Flu	ior® 633 A	Nexa Fluor [®] 64	47
Option for 3rd light source:									
violet diode laser 40 50 mW	05 nm		Alexa Fluor [®] 405						

Universidade de Coimbra
Rua Larga, Faculdade de Medicina,
Pólo I, 1º andarPólo III – Pólo das Ciências da Saúde
Universidade de Coimbra
Azinhaga de Santa Comba, Celas,UC – Biotech,
Parque Tecnológico de Cantanhede
Núcleo 04, Lote 8SectorNúcleo 04, Lote 85 3004-504 Coimbra, Portugal

T+351 239 820 190

F+351 239 822 776

3004-504 Coimbra, Portugal


T+351 239 480 200

Nucleo 04, Lote 8 5 3060-197 Cantanhede, Portugal T+351 231 249 170



Optical layout

The following figure shows a schematic diagram of a CyFlow[®] space optical layout, with detectors (FSC, SSC, FL1-7).



Electronics

Parallel real time signal processing for each of the optical channels with selectable linear, 3- or 4-decade logarithmic amplification (6-decade log amplifiers), pulse height, area and width analysis for doublet discrimination, 16 bit analog-to-digital converters, trigger to any parameter or all parameters.

The instrument is equipped with a CCD camera to monitor the particle flow and to check the focus of the objective.

Partec True Volumetric Absolute Counting

The Partec CyFlow[®] space allows the analysis of concentrations of samples by the method of True Volumetric Absolute Counting that is based on the precise measurement of a fixed sample volume by means of two electrodes.

> Universidade de Coimbra Rua Larga, Faculdade de Medicina, Universidade de Coimbra Pólo I, 1º andar 3004-504 Coimbra, Portugal

T+351 239 820 190

E+351 239 822 776


Pólo III – Pólo das Ciências da Saúde Azinhaga de Santa Comba, Celas, 3004-504 Coimbra, Portugal

T+351 239 480 200

UC - Biotech, Parque Tecnológico de Cantanhede Núcleo 04, Lote 8 3060-197 Cantanhede, Portugal T+351 231 249 170



The following figure shows a schematic representation of the True Volumetric Absolute Counting Electrode principle.



Partec Particle and Cell Sorter (PPCS)

The Partec particle and cell sorter, PPCS, serves as an optimized module for simplified and precise cell sorting. In order to overcome the problems of droplet sorters (biohazard, contamination) Partec has developed this closed sorting system. The PPCS sorting device can be sterilized safely. Any environmental contamination is completely avoided under normal operation conditions. In addition, the sorting action takes place smoothly. Fragile cells or large particles are sorted without any distortion and other forces acting on the cells during deflection. Long sorting runs (for several hundred milliliters of sample volume) are possible.

Sorting flow cell	Closed quartz type flow cell with 200 μm sorting channel for particle sizes up to 60 μm diameter
PPCS sorting control unit	Setting of sorting parameters for piezo crystal energy, delay, slope and hold time. Direct connection to CyFlow processor
Sorting speed	Input up to 15000 cells per second(maximum data acquisition rate), output for high purity (95%) of sorted fraction up to 300 cells per second. Highest purity (99%) with 100 cells per second
Sorting action	Piezo cristal for controlled particle deviation. Adjustable delay between point of analysis and deflection
Sort Collection Devices	One-way sorting to 3,5 ml, 55 × 12 mm tubes or to a sterile Shott flask

Universidade de Coimbra Rua Larga, Faculdade de Medicina, Pólo I, 1º andar 3004-504 Coimbra, Portugal

T+351 239 820 190

E+351 239 822 776

Pólo III – Pólo das Ciências da Saúde Universidade de Coimbra Azinhaga de Santa Comba, Celas, 3004-504 Coimbra, Portugal


T+351 239 480 200

UC – Biotech, Parque Tecnológico de Cantanhede Núcleo 04, Lote 8 **7** 3060-197 Cantanhede, Portugal T+351 231 249 170



Partec Particle and Cell Sorter Principle of Operation

The sorting flow cell has a y-shaped capillary channel system. Beyond the measuring area (laser focus/lamp focus) the flow channel is divided into the waste and sort channel, respectively. The piezo crystal is located in the waste channel and generates high speed pressure waves while cells are detected in the software sort regions. This pressure wave deflects all those cells out of the laminar flow into the sort channel outlet. The piezo can be driven several thousands times per second. The sample flow enters the sorter cuvette in a laminar stream, which is embedded in the sheath. This laminar flow is not disturbed during piezo action. Due to an exact timing condition the generated pressure wave acts only on the cell that travelled the illuminated light focal area. Any signal (laser forward or laser side scatter, laser or lamp induced fluorescence) or gate region is selectable for active sorting. The following figure shows a schematic representation of the sorter flow cell.



Universidade de CoimbraPólo III – Pólo das CiênciasRua Larga, Faculdade de Medicina,Universidade de CoimbraPólo I, 1º andarAzinhaga de Santa Comba,3004-504 Coimbra, Portugal3004-504 Coimbra, Portugal

T+351 239 820 190 F+351 239 822 776 Pólo III – Pólo das Ciências da SaúdeUC – Biotech,Universidade de CoimbraParque TecnolAzinhaga de Santa Comba, Celas,Núcleo 04, Lo3004-504 Coimbra, Portugal3060-197 Ca

T+351 239 480 200

UC – Biotech, Parque Tecnológico de Cantanhede Núcleo 04, Lote 8 **8** 3060-197 Cantanhede, Portugal info@cnc.uc.pt www.cnc.uc.pt

T+351 231 249 170



Flow Cytometry at CNC-UCBiotech:

Accuri™ C6 with autosampler (Becton Dickinson) – 4 colours Software: Accuri C6

Performance

Fluorescence Sensitivity	FITC < 150; PE < 100 MESF values determining using
	Thermo Scientific Cyto-Cal™ Multifluor Plus Violet Beads
Fluorescence Precision	<3% CV for CEN
Scatter Resolution	Scatter Resolution Resolves human peripheral blood
	lymphocytes, monocytes and granulocytes
Data Acquisition Rate	10,000 events/second, maximum

Fluidics

Flow Cell	200 µm ID quartz capillary
Minimum Detectable Particle Size	0.5 μm
Minimum Sample Volume	50 μL for tubes or plates
Pre-Set Flow Rates and Core Sizes	Slow: 14 μL/min, 10-μm core Medium: 35 μL/min, 16-μm core Fast: 66 μL/min, 22-μm core
Custom Sample Flow Rates	10-100 µL/min
Custom Core Diameter	5-40 μm
Recommended Sheath Fluid	0.2-µm filtered DI water
Maximum Events per Sample	1 million events per well

T+351 239 820 190

F+351 239 822 776

T+351 239 480 200

Universidade de CoimbraPólo III – Pólo das Ciências da SaúdeUC – Biotech,Rua Larga, Faculdade de Medicina,
Pólo I, 1º andarUniversidade de CoimbraPárque Tecnológico de Cantanhede3004-504 Coimbra, PortugalAzinhaga de Santa Comba, Celas,
3060-197 Cantanhede, PortugalNúcleo 04, Lote 89 T+351 231 249 170



Compatible Plate Types	Standard 96-well (flat, round, and v-bottom) plates
Processing Time	<90 minutes for 96-well plate, utilizing 30-second acquisition time per well

Laser Excitation	488 nm and 640 nm
Laser Profile	$10 \times 75 \ \mu m$
Light Scatter	Forward (0°, ±13°)
Detection	Side (90°, ±13°)
Emission Detection	4 colors, user-changeable optical filters
	Standard set installed:
	• FL1 533/30 nm (eg, FITC/GFP)
	• FL2 585/40 nm (eg, PE/PI)
	• FL3 > 670 nm (eg, PerCP, PerCP-Cy5.5, PE-Cy™7)
	• FL4 675/25 nm (eg, APC)
Optical Alignment	Fixed Alignment

Laser	Detector	Filter (nm)	Fluorochrome
Blue (488 nm)	FL1	530/30	FITC, GFP, CFSE, BB515
	FL2	585/40	PE , PI, PE-Texas Red
	FL3	670 LP	PerCP-Cy5.5, PE-Cy5.5, PE-Cy7, 7-AAD
	FL3	675/25	PerCP-Cy5.5, PE-Cy5, 7-AAD
	FL4	675/25	PerCP-Cy5.5, PE-Cy5
	FL3	610/20	PI, PE-Texas Red

Universidade de Coimbra Rua Larga, Faculdade de Medicina, Pólo I, 1º andar 3004-504 Coimbra, Portugal

T+351 239 820 190

F+351 239 822 776

Pólo III – Pólo das Ciências da Saúde Universidade de Coimbra Azinhaga de Santa Comba, Celas, 3004-504 Coimbra, Portugal

T+351 239 480 200

UC – Biotech, Parque Tecnológico de Cantanhede Núcleo 04, Lote 8 10 3060-197 Cantanhede, Portugal T+351 231 249 170



	FL4	610/20	PI
	FL3	630/30	7-AAD
	FL4	630/30	PE-Texas Red
	FL3	780/60	PE-Cy7
Red (648nm)	FL4	675/25	APC, Alexa-647, PE-Cy5
	FL3	780/60	APC-Cy7

FACSAria™ III (Becton Dickinson) – 12 colours Software: BD FACSDiva™

Performance

Fluorescence	FITC < 85 MESF values; PE < 29 MESF measurements
Sensitivity	performed at 70 psi and 90 kHz using SPHERO™ Rainbow
	Calibration Particles (RCP-30-5A)
Fluorescence	CV <3.0%, full G ₀ /G ₁ peak (area) for propidium iodide (PI)-
Resolution	stained chicken erythrocyte nuclei (CEN)
Fluorescence	Doublet/singlet ratio: CEN stained with PI: 1.95-2.05 (488-
Linearity	nm laser)
Forward and Side	Sensitivity enables separation of fixed platelets from noise,
Scatter Sensitivity	identification of bacteria, and detection of 0.5-µm beads
Forward and Side	Scatter performance is optimized for resolving lymphocytes,
Scatter Resolution	monocytes, and granulocytes
Sample	Maximum acquisition rate with 12 compensation pairs and 8
Acquisition Rate	parameters: 10,000 events/second

Fluidics

Sample Flow Rates	Adjustable dynamic range of sample flow rates
Fluidic Cleaning Modes Included (Software)	 Automated start up and shutdown Clean flow cell Prepare for aseptic sort

3004-504 Coimbra, Portugal

T+351 239 820 190

F+351 239 822 776

3004-504 Coimbra, Portugal

T+351 239 480 200

Universidade de CoimbraPólo III – Pólo das Ciências da SaúdeUC – Biotech,
Parque Tecnológico de CantanhedeRua Larga, Faculdade de Medicina,
Pólo I, 1º andarAzinhaga de Santa Comba, Celas,
Azinhaga de Santa Comba, Celas,UC – Biotech,
Parque Tecnológico de Cantanhede 3060-197 Cantanhede, Portugal T+351 231 249 170



Nozzles	70, 85, 100 and 130-µm, removable and can be sonicated. A registered key-fit position at the bottom of the cuvette provides fixed stream alignment
Automated Cell	ACDU for slide and plate sorting: 6, 24, 48, 96 and 384-
Deposition Unit	well plates

Laser Excitation	488 nm, 405 nm and 640 nm
Forward Scatter	Photodiode with 488/10 bandpass filter for the 488-nm
Detector and Filters	laser
Side Scatter Detector	Photomultipler with a 488/10 bandpass filter for the 488-
	nm laser

Laser	Detector	Mirror	Filter	Fluorochrome
Blue (488 nm)	А	750 LP	780/60	PE-Cy7
			BP	
	В	635 LP	695/40	PerCP/PerCP-Cy5-5/ 7-AAD/PE-
			BP	Cy5
	С	595 LP	616/23	PE-Texas Red/PE-CF594
			BP	
	D	550 LP	582/15	PE
			BP	
	E	495 LP	530/30	FITC/GFP/Alexa488/BB515/CSFSE
			BP	
	F		488/10	SSC
			BP	
Violet (405nm)	A	735 LP	780/60	Brilliant Violet 786
	В	685 LP	710/50	Brilliant Violet 711
	С	630 LP	660/20	Brilliant Violet 650
	D	600 LP	610/20	Brilliant Violet 605
	E	505 LP	525/50	Brilliant Violet 510/ AmCyan/CFP/
			BP	Pac Orange/V500
	F		450/50	Brilliant Violet 421/ Pacific
			BP	Blue/Cascade blue/V450/VPD450
Red (640 nm)	A	750 LP	780/60	APC-Cy7/APC-H7
			BP	

Universidade de Coimbra Rua Larga, Faculdade de Medicina, Pólo I, 1º andar 3004-504 Coimbra, Portugal

T+351 239 820 190

F+351 239 822 776

Pólo III – Pólo das Ciências da Saúde Universidade de Coimbra Azinhaga de Santa Comba, Celas, 3004-504 Coimbra, Portugal

T+351 239 480 200

UC – Biotech, Parque Tecnológico de Cantanhede Núcleo 04, Lote 8 <u>12</u> 3060-197 Cantanhede, Portugal T+351 231 249 170



В	685 LP	730/45 BP	Alexa-700
С		670/30 BP	APC/Alexa647

Sort Performance

Drop Drive Frequency	Range 1-100,000 Hz
Purity and Yield	At 70 psi and 90 kHz with an average threshold rate of 25,000 events per second, a four-way sort achieved a purity of >98% and a yield >80% of Poissonn's expected yield
Sort Collection Devices	Two-way sorting: 12 x 75 mm, and 15 mL Four-way sorting: 1.5 mL microtube and 12 x 75 mm
BD FACS [™] Accudrop	Red diode laser provided for fully automated drop-delay determination; Automated drop breakoff monitoring; Automated clog detection and sort tube protection system using Sweet Spot technology

Signal Processing

Converter	10-MHz Analog-to-Digital converter
Workstation	262,144 channels
Resolution	
Pulse Processing	Height, Area, and Width measurements available for any parameter. Ratio measurements are also available
Time	Time can be correlated to any parameter for kinetics experiments or other applications
Channel Threshold	Available for any parameter from any lasers with the ability to use multiple thresholds from different lasers simultaneously

Equipment acquisition was financed through the QREN - Centro Regional Operational Program 2007-2013 with the support of Mais Centro and the European Union under the project "CNC Biotech - Biotechnology research and business sector training".

3004-504 Coimbra, Portugal

T+351 239 820 190

F+351 239 822 776

3004-504 Coimbra, Portugal

T+351 239 480 200

Universidade de CoimbraPólo III – Pólo das Ciências da SaúdeUC – Biotech,Rua Larga, Faculdade de Medicina,Universidade de CoimbraParque Tecnológico de CantanhedePólo I, 1º andarAzinhaga de Santa Comba, Celas,Núcleo 04, Lote 82004 504 CaimbraDartural2004 504 Caimbra Núcleo 04, Lote 8 13 3060-197 Cantanhede, Portugal T+351 231 249 170